The single most important criterion for selecting medical gloves:

Ability to provide very effective barrier performance against blood pathogens and harmful infections.

# **Studies on Comparative Barrier Performance of Gloves**

Author	Failure Rates (%)		
	NATURAL RUBBER	VINYL	
Korniewicz 1990	7	63	
Korniewicz 2002	2.2	8.2	
Klein 1990	<1 <sup>a</sup>	22 <sup>a</sup>	
	<1 <sup>b</sup>	56 <sup>b</sup>	
Olsen 1993	4.2 – 7.9 <sup>c</sup>	43	
Douglas 1997	1.1 <sup>c</sup>	25 – 32 <sup>d</sup>	
		22 – 27 <sup>e</sup>	
Rego 1997	0-4	26 – 61 <sup>d</sup>	
		12 – 20 <sup>e</sup>	
Kerr (FDA) 2004	4 –10 <sup>c</sup>	38 <sup>c</sup>	
	9 –17 <sup>f</sup>	33 <sup>f</sup>	

a:No alcohol Pretreatment

d: Standard vinyl

b: Pretreatment with 70% alcohol

e: Stretch vinyl

c:NR latex powder-free

f: Powdered gloves

### MALAYSIAN RUBBER EXPORT PROMOTION COUNCIL

**HEAD OFFICE** Level 36, Q Sentral 2A Jalan Stesen Sentral 2. KL Sentral, 50470

60 3 2782 2100

3516, International Court N.W Washington DC 20008 USA.

Tel : +1 202 572 9771/9721 : +1 202 572 9787 Fax

CHINA OFFICE (SHANGHAI) Unit 1708, Hong Kong Plaza, 283 Huaihai Road Central, Huangpu District,

+86 21 3376 7001

CHINA OFFICE (GUANGZHOU) INDIA OFFICE

Room 505, Block A, China Shine Plaza. No.9 Linhe Xi Road, Tianhe,

907, Tulsiani Chambers Nariman Point Mumbai-400 021

: +91 22 6221 6725 : +91 22 6223 2359

**EUROPE OFFICE** : europe@mrepc.com **TYPE I** 

**STANDARD** MALAYSIAN GLOVE

Certified **Quality Glove** 

**Consistent Quality** 

Minimal **Protein Content** 

**Environment** Friendly



**Excellent Barrier Performance** 

> Unequaled Comfort, Fit and Feel

**High Strength** and Elasticity



SMG - Your Choice for Quality

www.smgonline.biz





This information is assembled and distributed by the Malaysian Rubber Export Promotion Council (USA), which has been registered under the Foreign Agents Registration Act as an agent of the Malaysian Rubber Export Promotion Council (Malaysia). Public information about this registration is available at the U.S. Department of Justice.

# **The Facts About Standard Malaysian Glove (SMG)**

### High in Barrier Performance and Low in Protein and Powder

The Standard Malaysian Glove (SMG) product quality certification scheme for latex examination gloves has been developed by Malaysia in consultation with various relevant authorities including the U.S. Food and Drug Administration (FDA). Latex gloves with SMG certification, lightly powdered and powder-free, meet stringent requirements for safety and reliability equivalent to international standards and are manufactured at facilities that comply with established quality management system. The SMG scheme establishes standards not only for barrier performance, the single most important function of medical gloves, but also for protein and powder content, elements believed to cause allergic reactions in individuals sensitive to latex proteins.

SMG-certified gloves can be identified by a logo, in green for powder-free examination gloves and in orange for lightly powdered examination gloves. The logos may also be in the dominant color of the dispenser box.





### **Environment Friendly Choice**

#### Renewable and Sustainable

SMG gloves are made from natural rubber latex, the sap of the Hevea brasiliensis tree widely grown in Malaysia. Bred to be the most efficient source of rubber latex, rubber trees are a renewable and sustainable resource.

#### Aid to Atmospheric Renewal

Studies estimate that rubber trees planted worldwide replace as much as 363,000 tonnes of Carbon Dioxide with 264,000 tonnes of oxygen annually through photosynthesis, revitalizing the earth's atmosphere and helping to combat greenhouse emissions and global warming.

#### Biodegradable

Natural rubber and its products are biodegradable, unlike synthetic rubber and plastics, such as polyvinyl chloride (PVC).

# **Ensuring Quality Management**

SMG manufacturers are required to implement an established quality management system to ensure continuous compliance with the standard or specification. SMG certification is carried out by the Malaysian Rubber Board (MRB), an internationally recognized organization that has been conducting research on rubber and rubber products for more than 80 years. MRB is accredited to ISO/IEC Guide 65 (general requirement for bodies operating product certification system). An independent Quality Inspectorate of the MRB undertakes regular surveillance testing to ensure quality compliance by SMG manufacturers. The MRB Laboratory is accredited to ISO/IEC 17025 (general requirements for the competence of testing and calibration laboratories) and is fully equipped for testing SMG gloves.

### **Ensuring Barrier Performance**

All SMG-labeled gloves must meet the prescribed standard for the water leak test, the internationally recognized testing procedure for barrier performance. The acceptable quality level (AQL) for SMG-certified gloves for this test must not exceed 1.5, or have more than 1.5 percent defectives, compared to the current higher level of 2.5 permitted under ASTM and FDA requirements for examination gloves. SMG gloves are sampled and inspected in accordance with ISO 2589-1 (Sampling procedures for inspection by attributes - Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection).

High strength is important to minimize the possibility that gloves will rupture or tear under stress during use. This characteristic is measured on the basis of tensile strength and elongation-at-break. The requirements under the SMG program are as shown below:

		Minimum Requirements			
		SMG	ASTM	D3578	EN455
Property	Unit	Type I	Type I	Type II	
1. Water leak test	AQL	1.5	2.5	2.5	1.5
2. Tensile strength before accelerated ageing	MPa	18	18	14	-
3. Elongation-at-break before accelerated ageing	%	650	650	650	-
4. Force-at-break before accelerated ageing	N	-	-	-	6
5. Tensile strength after accelerated ageing	MPa	14	14	14	-
6. Elongation-at-break after accelerated	%	500	500	500	_
7. Force-at-break after accelerated ageing	%	-	-	-	6

Note: Property No.2 to 7: Higher number denotes higher strength

SMG Type I: Conforms to USFDA requirements

ASTM Type I: Gloves with maximum stress at 500% elongation of 5.5 MPa

ASTM Type II: Gloves with maximum stress at 500% elongation of 2.8 MPa

#### Ref.

MRB's SMG Technical Requirements for Standard Malaysian Gloves, 3rd Revision Publication 4

ASTM D3578 – 05 (Reapproved 2010): Standard Specification for Rubber Examination Gloves

EN 455-1: 2000 Medical Gloves for Single Use-Requirements and Testing for Freedom from Holes

EN 455-2:2009+A1:2011 Medical Gloves for Single Use-Requirements and Testing for Physical Properties

## Minimizing Protein Sensitization

Research shows that latex protein allergies arising from the presence of high levels of residual extractable proteins, especially in highly powdered gloves, can be alleviated through the use of low-protein gloves. The SMG scheme is designed to ensure that natural rubber latex gloves consistently meet the low protein and powder contents as stipulated in the SMG standards.

In the case of SMG powdered gloves, protein and powder contents are kept at very low levels of 200  $\mu$ g /dm² and 150 mg per glove or less respectively. The permitted protein and powder contents of SMG powder-free gloves are even lower at 50  $\mu$ g /dm² and 2 mg per glove. The table below shows the upper limits set for protein and powder under the SMG scheme.

Property	SMG	ASTM D3578	EN455
Powder-free Upper limit of protein Upper limit of powder	50 μg/dm <sup>2</sup> 2 mg/glove	200 μg/dm <sup>2</sup> 2 mg/glove	2 mg/ glove
Powdered Upper limit of protein Upper limit of powder	200 μg/dm² 150 mg/glove (Approx 14 mg/ dm²)	200 μg/dm² 10 mg/glove	> 2 mg/ glove

#### Ref.

ASTM D6124 - 06(2011) Standard Test Method for Residual Powder on Medical Gloves EN 455-3:2006 Medical Gloves for Single Use: Requirements and Testing for Biological Evaluation